



Update-Description

inLab CAD 19.2

This document describes changes and improvements compared to previous version inLab CAD 19.1

February 2020

Following topics have been corrected

- If the machine is subsequently changed in administration from inLab MC XL to inLab MC X5, the insertion axis is retained and is not discarded.
- The initial proposal for restorations with "copy and mirror" were sometimes bad. The reference tooth is considered better now to improve the proposal.
- Optimizations for attachments in the "Finalize" step have been made.
- The calculation of the geometry of pontics has been improved.
- Corrections in the selection of implant systems for prefaces have been made. (Conelog and Xive)
- In some cases there were errors in the case preview, this behavior has been corrected.
- Virtual seating did not work properly in some situations after saving the case. This behavior has been corrected.
- Versions of license manager and portal client have been updated.

Quality improvement and stability

Dentsply Sirona continuously improves the quality of its products. As part of this release, several improvements have been made and known software bugs have been fixed. We always recommend using the latest software version.